

A

B

C

D

Mixer Located Below View Cell

Fiber Optic Illuminator/  
Quartz Windows

Vent Valve & Sample Scoop

Cabinet Enclosure

Color CCD Video Camera

Quartz Window

RTD Temperature Sensor


10,000 PSI Pressure Transducer

10,000 PSI Rupture Disc Assembly

CO<sub>2</sub> Liquid Exhaust

CO<sub>2</sub> Liquid Inlet

THIS DRAWING MAY CONTAIN CONFIDENTIAL MATERIAL AND IS THE PROPERTY OF SUPERCRITICAL FLUID TECHNOLOGIES, INC. IT SHOULD NOT BE DISCLOSED TO ANYONE WITHOUT EXPRESS WRITTEN PERMISSION FROM SFT, INC.



(P) 302-738-3420  
(F) 302-738-4320  
ken.james@supercriticalfluids.com

CLIENT

**SUPERCritical FLUID TECHNOLOGIES, INC.**

TOLERANCES (EXCEPT AS NOTED)

DECIMAL  
3 PL ± .005  
2 PL ± .01

FRACTIONAL  
± 1/16

ANGULAR  
± 0° 30'

MACHINE

SFT Phase Monitor

TITLE

Flow Diagram

ENGINEER

WCM

DRAWN BY

WCM

CHECKED BY

REV

A

SCALE

DRAWING NO.

SHEET NO.

1 of 1

A

B

C

D